

RESEARCH

Open Access



A novel magnet based 3D printed marker wand as basis for repeated in-shoe multi segment foot analysis: a proof of concept

Maarten Eerdeken¹, Filip Staes², Thomas Pilkington³ and Kevin Deschamps^{2,4,5*}

Abstract

Background: Application of in-shoe multi-segment foot kinematic analyses currently faces a number of challenges, including: (i) the difficulty to apply regular markers onto the skin, (ii) the necessity for an adequate shoe which fits various foot morphologies and (iii) the need for adequate repeatability throughout a repeated measure condition. The aim of this study therefore was to design novel magnet based 3D printed markers for repeated in-shoe measurements while using accordingly adapted modified shoes for a specific multi-segment foot model.

Methods: Multi-segment foot kinematics of ten participants were recorded and kinematics of hindfoot, midfoot and forefoot were calculated. Dynamic trials were conducted to check for intra and inter-session repeatability when combining novel markers and modified shoes in a repeated measures design. Intraclass correlation coefficients were calculated to determine reliability.

Results: Both repeatability and reliability were proven to be good to excellent with maximum joint angle deviations of 1.11° for intra-session variability and 1.29° for same-day inter-session variability respectively and ICC values of >0.91.

Conclusion: The novel markers can be reliably used in future research settings using in-shoe multi-segment foot kinematic analyses with multiple shod conditions.

Keywords: Marker, Magnet, In-shoe, Multi-segment, Gait, 3D

Background

Multi-segment foot kinematic analyses have gained enormous popularity in the last ten years since they provide a more detailed analysis of foot kinematics compared to single segment models [1, 2]. Until now, these multi-segment approaches have predominantly been used to assess barefoot foot kinematics [3–5]. Non-invasive in-shoe foot kinematics however, remain delicate to quantify as certain challenges arise: the difficulty to apply regular markers due to the fact that the skin is not directly accessible, the necessity for an adequate shoe which fits various foot morphologies and the need for adequate repeatability throughout a repeated measure condition. Several authors

tried to overcome these challenges by introducing modified shoes that allow skin-mounted markers for in-shoe multi-segment foot analysis with and without foot orthoses [6–10]. Apart from shoe modification, other challenges occur when measuring in-shoe foot kinematics, as regular markers are complex to place on the foot when in shod condition. Bishop et al. have addressed this challenge by manufacturing a so-called two-part in-shoe marker wand [7]. Being an interesting development, our aim was to expand the applicability of this concept in order to provide a solution for future research investigating shod conditions/orthoses, as an optimal combination between modified shoes and feasible markers needs to be found for repeated measurement designs. The objective of the current study was to develop and assess reliability and repeatability of a thin and solid in-shoe marker wand, consisting of a baseplate and marker-unit, with user-friendly features that allow for repeated in-shoe multi-segment foot analyses using modified shoes.

* Correspondence: kevin.deschamps@kuleuven.be

²Department of Rehabilitation Sciences– Musculoskeletal Rehabilitation Research Group, KULeuven, Tervuursevest 101, B-3001 Leuven, Heverlee, Belgium

⁴Institut D'Enseignement Supérieur Parnasse Deux-Alice, Division of Podiatry, Bruxelles, Belgium

Full list of author information is available at the end of the article



Methods

Participants

Ten asymptomatic adults signed the informed consent and were selected on their anthropometric heterogeneity (Table 1). The Ethical committee of University Hospitals of Leuven (S57147) granted approval for the study. Exclusion criteria were any trauma possibly affecting normal gait. Sample size estimation (Power: 0.80, type-1 error: 0.05) for ICC values was conducted and suggested that 13 participants suffice for this reliability study when an expected reliability (ρ_1) of 0.90 was applied [11].

In-shoe wand marker and shoe modification

The newly developed magnet-based in-shoe wand marker consists of two units, where one unit functions as the baseplate and the second as a rod, containing a retro reflective sphere (Fig. 1). The baseplate (length 18.0 mm, width 13.0 mm and height 3.00 mm) has five 0.80 mm holes and one 5.00 mm hole, with a disc magnet glued in the latter. The five 0.80 mm holes are made with a same inter-hole distance around the magnet-hole and provide additional support to the rod. The marker rod (stick height 5.00 mm, stick width 4.00 mm, sphere diameter 9.00 mm and total height 15.5 mm) possesses a magnet fixed in the stick to connect with the magnet of the baseplate. As mentioned, two disc magnets are used, one for the baseplate (diameter 5 mm, height 2 mm) and one for the marker rod (diameter 3 mm, height 3 mm), both with a force equal to 5.1 and 2.84 Newton respectively (Supermagnete, Webcraft GmbH, Gottmadingen, Germany). Since accelerations during walking remain low, orthogonal forces applied onto the magnets are inferior to the magnet forces, causing that the marker wands stay fixated into their baseplates throughout measurements. The structure component of both units was 3D printed in VeroWhitePlus Polyjet



Fig. 1 Novel magnet based in-shoe marker. Top: novel designed magnet based 3D printed marker, consisting of the baseplate (left) and the retro-reflective sphere unit (right). Bottom left: markers placed barefoot according to Rizzoli foot model. Bottom right: markers placed on foot within modified shoes for in-shoe multi segment foot analysis using the Rizzoli foot model

Table 1 Demographic and anthropometric data of recruited participants

Participant/ Gender/Age	Height (m)	Body mass (kg)	Shoe size (European)	FPI	Navicular height (mm)
1/M/36	1.80	72	43 1/2	5	42.1
2/M/28	1.79	90	44	2	62.1
3/F/25	1.76	71.8	42	2	51.3
4/F/24	1.59	63.6	38	6	47.9
5/M/24	1.80	65.5	43 1/2	3	58.6
6/M/21	1.70	62.2	43	3	59.8
7/M/49	1.69	67.8	43	4	42.6
8/M/23	1.84	85	43 1/2	6	41.3
9/F/39	1.65	61	42	7	39.8
10/F/39	1.73	59	42	7	34.2

M Male, F Female, FPI foot posture index, FPI ranges from -12 (highly supinated) to +10 (highly pronated)

(Stratasys, Eden Prairie, Minnesota, US) as this material has a tensile strength of 50–65 MPa, a polymerized density of 1.17 g/cm³ and flexural strength of 110 MPa, which compared to other commonly used materials is high.

A closed-toe shoe (Gel-odyssey WR, ASICS) with a black nubuck upper, a conventional laced fastening consisting of 5 eyelets, a rigid heel counter, a pitch of 1.4 cm, a single density midsole, moderate stiffness of the forefoot, a rubber outsole, a forefoot sole flexion point at the level of the metatarsal heads and a rigid midfoot sole (sagittal and frontal plane) was used in current study.

The shoe was modified by manually cutting eight windows into the upper of the shoe. In order to accommodate all foot types and individual anatomical differences, windows with a diagonal ranging from 3.5 cm to 8 cm were applied. Dimensions of each of these windows depended on the anatomical site of the foot, e.g. navicular tuberosity and base of first metatarsal bone are nearby each other. Therefore, one larger medial window was cut in order to more easily palpate both landmarks. Windows were also cut according to shoe size. By stitching the edges, the remaining materials surrounding the

windows were subsequently reinforced. European shoe sizes of the modified shoes were unisex and ranged from 36 to 48 (Fig. 1).

Data collection

After clinical exam assessing the Foot Posture Index (FPI) [12], participants were instrumented with baseplates of the novel in-shoe wand marker according to the Rizzoli foot model [5] and these were fixated onto the skin using double-sided tape (Scotch® 19 mm Double Sided, 3 M, Minnesota, US). First, 5 static trials were captured in sitting position while baseplates were applied and remained in situ for the duration of the testing session. In between the 5 static trials, the therapist removed and reapplied the rod to the baseplates to assess repeatability and robustness of the magnet fitting. A sitting position was used as this facilitated the participants in keeping their foot as still as possible in between trials. Next, participants entered in shod condition, rods were attached on their baseplates and a static trial (stance) was captured. Then, per three representative walking trials, rods were detached from the baseplates and participants were asked to remove the shoes while baseplates remained on the foot. After a 1-min break, participants re-entered in shod condition, rods were replaced on the baseplates and dynamic trial capturing restarted. This process was repeated up to three times so that a total of 9 trials were captured with replacing of the marker wands after trial 3 and 6. Each trial was time-normalized by interpolating the gait cycle into 100 frames.

Marker trajectories were captured using ten T-10 cameras and Nexus software (Vicon Motion Systems Ltd., Oxford Metrics, Oxford, England).

Data analysis

A mean placement error for all five sitting static trials was measured to assess the robustness of the interference magnet fitting, using the standard deviations between each trial and compared these to the movement of a reference marker (lateral malleolus marker). Comparison with the reference marker was done in order to compensate for the involuntary movement of the foot and so solely assess marker placement errors of the novel magnet fitted marker wands.

Intra-session and same-day inter-session repeatability of all three walking sessions were calculated using standard deviations between trials within a session (intra-session) and between sessions (same-day inter-session). Standard deviations were assessed per frame of the gait cycle (100 frames) and a mean of all these standard deviations was calculated for easier comparisons. Ratio between inter- (nominator) and intra-session repeatability (denominator) was computed as well.

Intraclass correlation coefficients ($ICC_{3,1}$) between three walking sessions were calculated for three variables (maximum peak value, minimum peak value and range of motion) during gait cycle in order to assess reliability. Additionally, standard error of the mean (SEM) was calculated for all joint angle rotations and 95% confidence intervals of absolute ROM (range of motion) of each joint angle and plane were assessed as well. Statistics were computed using MedCalc (MedCalc software BVBA, Ostend, Belgium) and Excel (Microsoft Corporation, Redmond, USA).

Results

Mean placement error of all markers was 0.44 mm, with a maximum placement error of 1.16 mm and a minimum placement error of 0.2 mm (Fig. 2). Intra-session repeatability or natural gait variability varied between 0.30 degrees for calcaneal-midfoot joint in frontal plane (Y) and 1.11 degrees for shank-calcaneal joint in sagittal plane (X). For the inter-session repeatability, outcomes ranged between 0.41 degrees for calcaneal-midfoot joint in Y-plane and 1.29 degrees for midfoot-metatarsal joint in transverse plane (Z). Ratios mostly exceeded 1.00 (Table 2).

All ICCs were high (>0.91), except for calcaneus – midfoot rotation in transverse plane (0.792) (Table 3). SEMs remained below 0.8° for all joint rotation angles and 95% confidence interval, wherein the absolute range of motion is situated, was given as well (table 3).

Discussion

The aim of this study was to design a novel magnet based, 3D printed marker wand for in-shoe multi-segment foot analysis using a repeated measurement design. To gain confidence in using these markers in




Fig. 2 Mean marker placement errors. Mean placement errors of novel markers in sagittal plane (blue bar), frontal plane (orange bar) and transverse plane (grey bar). LCA: left calcaneus (reference marker); LPT: left peroneal tubercle; LST: left sustentaculum talus; LTN: left navicular tubercle; LVMB: left fifth metatarsal base; LSB: left second metatarsal base; LFMB: left first metatarsal base; LVM: left fifth metatarsal head; LSM: left second metatarsal head; LFM: left first metatarsal head; LPM: left first phalanx

Table 2 Intraclass correlation coefficients of joint kinematics

Joint rotation angle:	ICC (95% CI)			SEM (95% CI)
	Max	Min	ROM	ROM
Sha - Cal DF/PF	0.980 (0.942–0.995)	0.978 (0.937–0.994)	0.966 (0.904–0.991)	0.59 (24.8–27.1)
Sha - Cal Inv/Ev	0.991 (0.975–0.998)	0.993 (0.979–0.998)	0.948 (0.853–0.986)	0.34 (9.24–10.6)
Sha - Cal Add/Abd	0.993 (0.980–0.998)	0.997 (0.992–0.999)	0.948 (0.835–0.986)	0.38 (10.7–12.2)
Cal - Met DF/PF	0.995 (0.986–0.999)	0.993 (0.980–0.998)	0.960 (0.883–0.989)	0.54 (12.3–14.4)
Cal - Met Inv/Ev	0.988 (0.964–0.996)	0.979 (0.940–0.994)	0.947 (0.840–0.986)	0.29 (5.70–6.90)
Cal - Met Add/Abd	0.947 (0.843–0.986)	0.986 (0.960–0.996)	0.944 (0.834–0.985)	0.69 (6.63–9.34)
Cal - Mid DF/PF	0.980 (0.943–0.995)	0.979 (0.940–0.994)	0.968 (0.908–0.991)	0.21 (6.30–7.13)
Cal - Mid Inv/Ev	0.979 (0.943–0.995)	0.934 (0.808–0.982)	0.792 (0.338–0.954)	0.12 (2.78–3.27)
Cal - Mid Add/Abd	0.992 (0.978–0.997)	0.987 (0.963–0.996)	0.917 (0.754–0.977)	0.23 (5.51–6.41)
Mid - Met PF/DF	0.994 (0.984–0.998)	0.995 (0.984–0.998)	0.968 (0.909–0.991)	0.49 (9.59–11.5)
Mid - Met Inv/Ev	0.981 (0.947–0.995)	0.977 (0.934–0.994)	0.942 (0.830–0.984)	0.24 (2.73–3.69)
Mid - Met Add/Abd	0.965 (0.897–0.991)	0.985 (0.955–0.996)	0.969 (0.908–0.992)	0.72 (5.43–8.27)
Hallux DF/PF	0.991 (0.975–0.998)	0.991 (0.974–0.998)	0.914 (0.757–0.977)	0.29 (6.09–7.24)

Sha shank, Cal calcaneus Met metatarsal, Mid midfoot, PF plantar flexion, DF dorsiflexion, Max highest peak value of joint angle during gait cycle; Min, lowest peak value of joint angle during gait cycle; ROM, Range of motion during gait cycle; ICC, intraclass correlation coefficients; SEM standard error of the mean; 95% CI, 95% confidence interval; Joint angle rotation in degrees

research applications, repeatability and reliability was investigated for in-shoe walking trials.

Repeatedly taking on and off markers wands from the baseplates in static position generates low placement errors as the magnet fitting principle showed a solid robustness. This implies that once the baseplates are placed, marker wands can be taken on and off by any other therapist with the knowledge that outcomes will not vary significantly.

To assess repeatability when performing dynamic in-shoe multi-segment foot analysis with these new markers, we set up a design in which shoes ought to be taken on and off while baseplates remained on the foot during the whole measurements. Inter-session repeatability never exceeded 1.29 degrees (Table 3), meaning that this was the maximum difference in kinematic outcome values when measuring multiple shod conditions during one analysis. To correctly assess the influence of 1) natural gait variability and 2) repeatedly taking on and off the shoe, we calculated

the ratio between these two test conditions. Putting on shoes multiple times influenced kinematic waveforms more than natural gait variability did. Though, absolute error values still remained very low as previously mentioned. This demonstrates the advantage of the magnet fitting principle, which insures near to identical positions of the marker wand, even when they ought to be taken on and off their baseplates. Also, changing shoes will not alter the baseplate position, which initially was considered as a risk since some of the baseplates are located on protrudes anatomical entities.

Reliability of measuring similar in-shoe foot kinematics in a repeated-measurements design tended to be very high as suggested by the ICC outcomes. Looking at existing research regarding reliability of placing regular markers, these novel markers score higher on average as suggested by ICC outcomes compared to these other studies [13, 14]. Although, discretion in comparison is warranted as methodology and experiment set-up may differ between studies.

Table 3 Intra – and inter-session variability of kinematic measurements

Joint rotation angle:	Intra-session variability			Inter-session variability			Ratio (inter/intra)		
	X	Y	Z	X	Y	Z			
Sha - Cal	1.11	0.88	0.75	0.87	0.82	0.77	0.79	0.94	1.03
Cal-Mid	0.53	0.30	0.56	0.70	0.41	0.85	1.32	1.37	1.50
Cal-Met	0.74	0.49	0.55	0.93	0.68	1.25	1.25	1.38	2.26
Mid-Met	0.60	0.37	0.59	0.83	0.81	1.29	1.40	2.16	2.19
Hallux (planar)	-	-	0.48	-	-	0.97			2.00

Sha, shank, Cal calcaneus, Mid midfoot, Met metatarsal, X sagittal plane, Y frontal plane, Z transverse plane

All results in degrees, except for ratio

Assessment of Foot Posture Index was conducted to check heterogeneity in foot morphologies between participants and appeared relatively high (cfr. 3.1). For this, modification of the shoes was done by cutting larger holes to overcome issues occurring when multiple participants with heterogeneous feet ought to be tested with identical shoe sizes. This, however, is also a first limitation to current study, as some shoe integrity may be lost due to this modification. Yet, it would be costly and time consuming to modify a pair of shoe per participant. Another known limitation is the fact that experienced testers are still required, since placing of the baseplate determines outcome values. The reference marker used to account for foot movement is not ideal, since this would imply that ankle and foot are rigid. Yet, we were not able to add another reference marker onto the foot, since this was impossible to standardize in this type of measurement. Also, due to circumstances, we were not able to conduct an inter rater reliability analysis, which is a limitation to current study. However, we believe that potential inter rater reliability in these novel markers will be alike to earlier research studying marker placement reliability, as placing of the baseplates is similar to regular marker placement. Last, additional research will be necessary to gain confidence for use in a clinical population and also in settings where running trials need to be captured, since forces on the foot will be greater and reliability of the robustness of the magnet-based link needs to be re-evaluated.

Conclusion

The novel 3D-printed markers, consisting of a baseplate and a wand marker, are a reliable basis for future settings using in-shoe multi segment foot analysis when shoes ought to be taken on and off repeatedly. Results showed a solid robustness of the magnet fitting principle that allows for reliable kinematic outcomes in a repeated measurement design conducting in-shoe walking trials.

Abbreviations

FPI: Foot Posture Index; ICC: Intraclass correlation coefficient; RO: Range of motion; SEM: Standard error of the mean

Acknowledgements

None.

Funding

This study was supported by the ASPIRE 2015 (Pfizer) research grant (KD).

Availability of data and materials

The datasets used and analyzed during the current study are available from the corresponding author on reasonable request.

Authors' contributions

ME made substantial contributions to conception and design of the study, undertook all clinical data collection, contributed to analysis and interpretation of data and to writing the manuscript. FS made substantial contributions to interpretation of the data and writing the manuscript. TP made substantial contributions to conception and design of the study and

writing of the manuscript. KD made substantial contributions to conception and design of the study, undertook some clinical data collection, contributed to analysis and interpretation of data and to writing the manuscript. All authors were involved in drafting of the manuscript or revising it critically for important intellectual content. All authors read and approved the final manuscript.

Ethics approval and consent to participate

The study was approved by The Ethical committee of University Hospitals of Leuven (S57147). All participants provided written informed consent.

Consent for publication

Not applicable.

Competing interests

None of the authors of this study have any financial or personal relationships with other people or organizations that could inappropriately influence or bias their work.

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.

Author details

¹Clinical Motion Analysis Laboratory, UZ Pellenberg, Weligerveld 1, 3212 Lubbeek, Belgium. ²Department of Rehabilitation Sciences– Musculoskeletal Rehabilitation Research Group, KULeuven, Tervuursevest 101, B-3001 Leuven, Heverlee, Belgium. ³Fablab Leuven, Celestijnenlaan 300, 3001 Leuven, Heverlee, Belgium. ⁴Institut D'Enseignement Supérieur Parnasse Deux-Alice, Division of Podiatry, Bruxelles, Belgium. ⁵Department of Podiatry, Artevelde University College, Ghent, Belgium.

Received: 7 June 2017 Accepted: 8 August 2017

Published online: 18 August 2017

References

- Deschamps K, Staes F, Roosen P, Nobels F, Desloovere K, Bruyninckx H, Matricali GA. Body of evidence supporting the clinical use of 3D multisegment foot models: a systematic review. *Gait Posture*. 2011;33:338–49.
- Pothrat C, Authier G, Viehweger E, Berton E, Rao G. One- and multi-segment foot models lead to opposite results on ankle joint kinematics during gait: implications for clinical assessment. *Clin Biomech*. 2015;30:493–9.
- Bruening DA, Cooney KM, Buczek FL. Analysis of a kinetic multi-segment foot model. Part I: model repeatability and kinematic validity. *Gait Posture*. 2012;35:529–34.
- Saraswat P, MacWilliams BA, Davis RB, D'Astous JL. A multi-segment foot model based on anatomically registered technical coordinate systems: method repeatability and sensitivity in pediatric planovalgus feet. *Gait Posture*. 2013;37:121–5.
- Leardini A, Benedetti MG, Berti L, Bettinelli D, Nativio R, Giannini S. Rear-foot, mid-foot and fore-foot motion during the stance phase of gait. *Gait Posture*. 2007;25:453–62.
- Bishop C, Paul G, Thewlis D. The reliability, accuracy and minimal detectable difference of a multi-segment kinematic model of the foot-shoe complex. *Gait Posture*. 2013;37:552–7.
- Bishop C, Arnold JB, Fraysse F, Thewlis D. A method to investigate the effect of shoe-hole size on surface marker movement when describing in-shoe joint kinematics using a multi-segment foot model. *Gait Posture*. 2015;41:295–9.
- Shultz R, Jenkyn T. Determining the maximum diameter for holes in the shoe without compromising shoe integrity when using a multi-segment foot model. *Med Eng Phys*. 2012;34:118–22.
- Halstead J, Keenan AM, Chapman GJ, Redmond AC. The feasibility of a modified shoe for multi-segment foot motion analysis: a preliminary study. *J Foot Ankle Res*. 2016;9:7.
- Arnold JB, Bishop C. Quantifying foot kinematics inside athletic footwear: a review. *Footwear Sci*. 2013;5:55–62.
- Walter SD, Eliasziw M, Donner A. Sample size and optimal designs for reliability studies. *Stat Med*. 1998;17:101–10.

12. Redmond AC, Crosbie J, Ouvrier RA. Development and validation of a novel rating system for scoring standing foot posture: the foot posture index. *Clin Biomech.* 2006;21:89–98.
13. Deschamps K, Staes F, Bruyninckx H, Busschots E, Jaspers E, Atré A, Desloovere K. Repeatability in the assessment of multi-segment foot kinematics. *Gait Posture.* 2012;35:255–60.
14. Telfer S, Morlan G, Hyslop E, Semple R, Rafferty D, Woodburn J. A novel device for improving marker placement accuracy. *Gait Posture.* 2010;32: 536–9.

Submit your next manuscript to BioMed Central and we will help you at every step:

- We accept pre-submission inquiries
- Our selector tool helps you to find the most relevant journal
- We provide round the clock customer support
- Convenient online submission
- Thorough peer review
- Inclusion in PubMed and all major indexing services
- Maximum visibility for your research

Submit your manuscript at
www.biomedcentral.com/submit

